



	<b>Chemiluminescence Assay for Quantitative Determination of Anti-Mullerian Hormone (AMH) in Human Serum</b>
<b>Application</b>	For the quantitative determination of AMH in human serum.
<b>Principle</b>	Sandwich based Enzyme linked immunoassay (ELISA)
<b>Reagents</b>	<ul style="list-style-type: none"> <li>• <b>Coated microwells:</b> 96 microwells (3x8 wells x 4) are coated with monoclonal anti- AMH- antibody.</li> <li>• <b>Standards:</b> AMH standards with desired concentration for calibration in lyophilized form.</li> <li>• <b>AMH Enzyme Conjugate (100X)</b></li> <li>• <b>Sample Diluent.</b> Ready to use.</li> <li>• <b>AMH Conjugate Diluent.</b></li> <li>• <b>Substrate A:</b> Chemiluminescent substrate containing enhanced luminol solution.</li> <li>• <b>Substrate B:</b> Chemiluminescent substrate containing stabilized peroxide solution.</li> <li>• <b>Wash buffer:</b> Buffer containing surfactants (20X). To be diluted 20 times with distilled water or deionized water.</li> </ul>
<b>Test duration</b>	100 minutes
<b>Storage / Stability</b>	2-8°C, 12 months

Presentation	Pack	Cat.No.
<b>Electra AMH kit contains:</b> <ul style="list-style-type: none"> <li>• Microwell strips (3 x 8 wells/pouch)</li> <li>• Reagents</li> <li>• Accessories</li> <li>• Standards</li> </ul>	96 Tests	512070096