

	Chemiluminescence Assay for the Quantitative Determination of Free Prostate Specific Antigen (fPSA) in Human Serum.
Application	For the quantitative determination of Free Prostate Specific Antigen (fPSA) in Human Serum.
Principle	Chemiluminescence (CLIA)
Reagents	<ul style="list-style-type: none"> • Coated microwells: 96 microwells (3x8 wells x 4) are coated with Anti- fPSA antibody. • Standards: fPSA standards with desired concentration for calibration. • Enzyme Conjugate. Ready to use. • Sample Diluent: Ready to use. • Substrate A: Chemiluminescent substrate containing enhanced luminol solution. • Substrate B: Chemiluminescent substrate containing stabilized peroxide solution. • Wash buffer: Buffer containing surfactants (20X). To be diluted 20 times with distilled water or deionized water.
Test duration	40 minutes
Storage / Stability	2-8°C, 12 months

Presentation	Pack	Cat.No.
Electra fPSA kit contains: <ul style="list-style-type: none"> • Microwell strips (3 x 8 wells/pouch) • Reagents • Accessories • Standards 	96 Tests	513030096

